



MR Conditional Laryngoscope Set

SAFETY FIRST!
*Intubate confidently within
the MR environment*



- *Specially designed, lightweight low-magnetic signature Fibre Optic Laryngoscope System with brighter, whiter Xenon light.*
- *Distinctly Identifiable Gold and Black blade and handle design. Reduces risk of incompatible equipment getting into the MRI suite.*
- *Distinctly affordable Battery Pack. Reduces the drain on your budget and lowers costs of replacement parts*

Conforms to ISO 7376-1, ASTM F 965-99,
ASTM F 2052-02 and ASTM F 2503-05
Approvals - ISO 9001, ISO 13485
FDA registered manufacturer



TRUPHATEK[®]

Technical Specifications



Tru-MR™ Sets	
Codes	Description of contents
5442M	One handle, 5 blades (Miller 0 and 1, Mac 2, 3 and 4) and 2 Battery Packs
5443M	One handle, 6 blades (Mac 2,3 and 4, Miller 1,2 and 3) and 2 Battery Packs



Tru-MR™ FibreClips™	
Codes	Product Description
524102M	Mac 1 - Pack of 2
524202M	Mac 2 - Pack of 2
524302M	Mac 3 - Pack of 2
524402M	Mac 4 - Pack of 2
525002M	Miller 0 - Pack of 2
525102M	Miller 1 - Pack of 2
525202M	Miller 2 - Pack of 2
525302M	Miller 3 - Pack of 2
525402M	Miller 4 - Pack of 2



Tru-MR™ Handles	
Codes	Product Description
5525M	Medium Size



Tru-MR™ Accessories	
Codes	Product Description
53102M	2 Battery Packs

From Truphatek's Warranty Statement:

Truphatek® laryngoscope blades and handles carry a 5 year warranty from date of manufacture on all stainless steel parts against defects in manufacture and materials.

FibreClip™ replaceable fibre-optic light guides are guaranteed to perform for a period of 15 months from date of manufacture.

Tru-MR™ battery packs are guaranteed for 18 months from date of manufacture as appearing on the pack.

Full text of the Truphatek® warranty and further information about Truphatek® products are available on our website www.truphatek.com or through our Customer Service desk at helpdsk@truphatek.com.

Tru-MR™ products have been tested and demonstrated to be MR-Conditional by Shellock R&D Services, Inc., CA., USA (www.magneticresonancesafetytesting.com) for magnetic field interactions at 3-Tesla with highest spatial gradient field of 720-Gauss/cm or less.

Tru-MR™, FibreClip™ and Truphatek® are trademarks of Truphatek International Ltd. and may not be used without prior written permission of the owner.

Distributed by

TRUPHATEK®
www.truphatek.com

Truphatek International Ltd. P.O.Box 8051 Netanya 42504, Israel
Tel: 972-9-8851155, Fax: 972-9-8851212, helpdsk@truphatek.com

© 2006-2007 TRUPHATEK® All rights reserved

MR-1-04/07 Rev 2